

Short Communication

Development and Validation of a Complexometric Titration Method for the Determination of Rosuvastatin Calcium in Raw Material.

ABSTRACT

Aims: an easy, sensitive and inexpensive volumetric method for the determination of rosuvastatin calcium in raw material has been developed.

Methodology: The titrimetric method is based on the reaction of calcium with a solution of Disodium ethylene diaminetetraacetate (EDTA) - Magnesium 0.01M. Hydroxynaphtol blue was used as indicator. It changes from pink to blue at pH = 10 at the end point of the titration. The method was validated for linearity, precision and accuracy, following the suggestions of the International Conference on Harmonization (ICH) **Results:** The linearity of the volumetric method was determined by analysis of six replicates at 80%, 100% and 120% and three replicates at 90% and 110% of analyte concentration. The calibration curve was linear, with $r = 0.9998$. Assay method precision was evaluated by carrying out six independent assays of bulk drug and the intermediate precision was also verified using different analyst and different day in the same laboratory. Accuracy (mean recovery 99.0%) and precision were found to be satisfactory.

Conclusion: The proposed method can be used for quality control assay of rosuvastatin calcium in bulk drug.

Keywords: Rosuvastatin calcium , Titrimetric, Assay, Complexometric titration, EDTA.Mg.

1. INTRODUCTION

Rosuvastatin calcium (Fig. 1) is bis[(E)-7-[4-(4-fluorophenyl)-6-isopropyl-2- [methyl(methylsulfonyl)amino] pyrimidin-5-yl](3R,5S)-3,5-dihydroxyhept-6-enoic acid] calcium salt. It is a competitive inhibitor of the enzyme 3-hydroxy-3-methyl-glutaryl-CoA reductase (HMG-CoA reductase), the enzyme that converts 3-hydroxy-3-methylglutaryl coenzyme A to mevalonate, precursor for cholesterol. It is a cholesterol lower agent [1]. The usual doses are 5, 10, 20 and 40 mg.




Figure 1 Rosuvastatin calcium

Rosuvastatin calcium alone has been determined by Spectrophotometric methods [2-4], TLC methods [5-7] and a Stability indicating HPTLC method [8].


A literature study revealed some high-performance liquid chromatographic methods for the quantitation of rosuvastatin calcium [4, 9-13]. Most of the analytical techniques for rosuvastatin calcium described in the literature are based on the HPLC determination of this drug with another active drug substance [1, 14-22].

Due to the need for an absolute analytical technique for rosuvastatin calcium, we have developed this volumetric technique. The present technique is concerned using a complexometric titration for the

quantification of Calcium (Ca). Complexometric titrations with metal ions serve as analytical determinations of raw material at the millimole concentration. The substances that have a pair of unshared electrons (e.g. on N, O, S atoms in the molecule) are capable of satisfying the coordination number of the metal. The metal ion is a Lewis acid (electron pair acceptor) and the complexing agent is a Lewis base (electron pair donor). The complexing agent use a number of molecules which depends on the coordination number of the metal and the complexing groups of it.

A chelating agent is an organic substance that has two or more groups capable of complexing with a metal ion, forming a chelate. This titration is called a chelometric titration, which is a particular type of complexometric titration. There are six complexing groups in EDTA. The EDTA is represented by the symbol H₄Y. The four hydrogens in the formula refer to the four acidic hydrogens on the four-carboxyl groups. It is a triprotic acid. The pair of unshared electrons is located on each of the two nitrogen atoms and each of the four-carboxyl groups [23].

There are not satisfactory indicators for calcium by direct titration. Eriochrome Black T cannot be used as an indicator because it forms a weak complex with calcium to give a sharp end point. Nevertheless, with a small addition of magnesium chloride to the EDTA solution, calcium can be determined by direct titration. The titrant solution contains a magnesium-EDTA complex which is introduced into the titration mixture. As Ca²⁺ forms a more stable complex with EDTA than magnesium, the following reaction occurs:



Hydroxynaphthol blue is used as indicator, it is itself a chelating agent. It changes from pink to blue in the presence of calcium at pH 10.

The present manuscript describes an easy, sensitive and inexpensive volumetric method for the quantitation of rosuvastatin calcium in raw material.

The method was validated for linearity, precision and accuracy, following the suggestions of the International Conference on Harmonization (ICH) [24].

2. MATERIALS AND METHODS

2.1 Apparatus

A standard borosil burrets, pipetts, standard flasks, measuring cilindres and conical flasks are calibrated as per International Conference on Harmonization (ICH) guidelines [25]

2.2 Materials

Rosuvastatin calcium (100.0%) was obtained from Optimus drugs (Hyderabad, India). Disodium EDTA was AR Grade (Merck Química Argentina, Argentina); Magnesium chloride hexahydrate AR Grade (Alcor, Argentina) Ammonia AR Grade (Conc 20%, Reactivo Raudo Analítico, Argentina), Ammonium chloride AR Grade (J.T.Baker, Mexico), Calcium carbonate ACS grade (Mallinckrodt, USA), Hydrochloric acid ACS grade (J.T.Baker, Mexico), Hydroxy naphthol blue AR Grade (Anedra, Argentina). Purified (distilled) water was passed through a 0.45 µm membrane filter.

2.2.1 Magnesium chloride 1% (w/v)

1.07 g of magnesium chloride hexahydrate was taken in a 50 ml volumetric flask, dissolved with 30 ml of distilled water and made to volume with the same solvent.

2.2.2 Ammonia 6M

The 51 mL ammonia (NH₃) 20% was taken in 100 mL volumetric flask and made to volume with distilled water.

2.2.3 Disodium ethylene diaminetetraacetate (EDTA) - Magnesium Volumetric solution

4.0 g of disodium EDTA dehydrate salt was dissolved in 500 mL of distilled water in a 1 L volumetric flask. Add 10 mL of magnesium chloride 1% and 2 mL of ammonia 6M and made to 1 L with purified water.

2.2.4 Ammonia – ammonium chloride buffer

Dissolve 6.0 g ammonium chloride in 100 mL of ammonia 6M. Stir until total dissolution.

2.2.5 Diluted hydrochloric acid (0.1N)

Prepare by diluting 100 ml of 1N hydrochloric acid solution with sufficient water to make 1000 mL.

2.3 Standardization

The 0.01M EDTA - magnesium solution was standardized with calcium carbonate, placing 30 mg previously dried at 110 °C to constant mass in a 25.0 mL conical flask, dissolve in 10 mL of distilled water and stir until dissolution. Add 10 mL of 0.1N hydrochloric acid. Add 10 mL of ammonia – ammonium chloride buffer and 30 mL of distilled water. Add 25 mg of hydroxy naphthol blue and shake during 1 min. Until the pink colour changes to blue at the end of titration, the solution was titrated with 0.01M EDTA - magnesium solution. Each mL of 0.01 M EDTA - magnesium solution is equivalent to 1.0008 mg of calcium carbonate.

2.4 General Procedure

250 mg of rosuvastatin calcium was exactly weighed, placed in a 500 ml conical flask, and dissolved in 200 mL of purified water and 25 mL of ammonia – ammonium chloride buffer. Stir during 10 min. Add 25 mg of hydroxy naphthol blue and shake during 1 min (pH= 10.0) Titrate with 0.01M EDTA - magnesium solution until the pink colour changes to blue at the end of titration.

Each mL of 0.01 M EDTA - magnesium solution is equivalent to 10.0114 mg of rosuvastatin calcium.

2.5 Method Validation

2.5.1 Linearity

Linearity was determined by analysis of six replicates at 80%, 100% and 120% and three replicates at 90% and 110% of analyte concentration.

2.5.2 Precision

Precision of the assay method was performed by analyzing six individual samples of rosuvastatin calcium raw material. Intermediate precision was checked by analyzing six individual samples by two different analysts on different days.

2.5.3 Accuracy

Recovery studies were done to evaluate the accuracy of the method, 6 samples each at three different levels (80%, 100% and 120%). It was calculated the amount of rosuvastatin calcium recovered in relation to the added amount

3. RESULTS AND DISCUSSION

The described method has been suitably applied to the quantitation of rosuvastatin calcium in raw material.

Linearity was determined by analysis of six replicates at 80%, 100% and 120% and three replicates at 90% and 110% of analyte concentration. The correlation coefficient ("r") value was 0.9998 and the calibration curve showed adequate linearity over the concentration range. The regression equation for the calibration curve was found to be $y = 0.1035x + 0.2495$. The method was lineal, the calibration graphs was fulfilled by the high value of the correlation coefficient and the intercept value that was not statistically (P = .05) different from zero (Table 1 and Figure 2).

136
137
138
139

TABLE 1 Linearity

| % of nominal value | Weighed (mg) (RSD) | Vol cons (RSD) |
|--------------------|--------------------|----------------|
| 80 | 200.6 (0.6) | 21.0 (0.7) |
| 90 | 226.4 (0.5) | 23.7 (0.4) |
| 100 | 251.8 (0.6) | 26.3 (0.7) |
| 110 | 276.2 (0.1) | 28.7 (0.3) |
| 120 | 301.7 (0.5) | 31.5 (0.4) |

| | |
|--------------------------|----------------------|
| Slope ^(a) | 0.1035. \pm 0.0011 |
| Intercept ^(b) | 0.2495 \pm 0.3706 |

^aConfidence limits of the slope (p= 0.05)
^bConfidence limits of the intercept (p= 0.05)




Figure 2 Linearity

The precision is usually expressed as the RSD of a series of measurements. It expresses the proximity of compliance between a series of assessments obtained from multiple sampling of the aforementioned homogeneous specimen under the prescribed conditions.

The intermediate precision of the volumetric method was performed by analyzing the samples on two different days by two different persons. The results were presented both separately and as the mean. For precision tests the conclusion was: average values 99.5 and 100.9 %, RSD 0.4% and 0.5%. "t" Test comparing two assays with 95% confidence for 10 degrees of freedom disclosed that both means were not sufficiently different *inter se* ($t_{n-2, \alpha:0.05} = 2.23$ (Table 2).

TABLE 2. Intermediate precision

| Analyst 1 Sample N° | Weighed (mg) | Percentage | Analyst 2 Sample N° | Weighed (mg) | Percentage |
|------------------------|--------------|------------|------------------------|--------------|------------|
| 1 | 249.8 | 99.3 | 1 | 250.6 | 100.9 |
| 2 | 250.2 | 99.1 | 2 | 252.3 | 101.3 |
| 3 | 250.4 | 99.4 | 3 | 250.1 | 101.1 |
| 4 | 249.7 | 100.1 | 4 | 252.0 | 101.4 |
| 5 | 250.8 | 99.6 | 5 | 250.0 | 100.3 |
| 6 | 251.8 | 99.3 | 6 | 250.9 | 100.4 |
| Mean | 250.4 | 99.5 | | 251.0 | 100.9 |
| RSD | 0.3 | 0.4 | | 0.4 | 0.5 |

The percentage recuperation range and RSD values were 98.5-99.7% and within 0.3% respectively. Accuracy was also evidenced by making a graph of the amount of rosuvastatin calcium measured against the amount present in the samples, both expressed in mg. Linear regression analysis provided slopes not sufficiently different from 1 (t test $P=0.05$), intercepts not significantly different from zero (t test $P=0.05$) and $r = 0.9999$, the RSD was 0.3. (Table 3 and Figure 2)

201 **TABLE 3. Accuracy**
202

| % of nominal value | Added amount (mg) | Found amount (mg) | Recovery (%) | Average recovery (n=3) | RSD (%) |
|--------------------|-------------------|-------------------|--------------|------------------------|---------|
| 80 | 194.0 | 191.6 | 98.8 | 99.1 | 0.4 |
| | 194.2 | 192.5 | 99.1 | | |
| | 193.1 | 190.7 | 98.8 | | |
| | 192.2 | 189.8 | 98.7 | | |
| | 193.1 | 191.6 | 99.2 | | |
| | 193.0 | 191.6 | 99.3 | | |
| | 194.0 | 193.5 | 99.7 | | |
| 100 | 241.2 | 238.2 | 98.8 | 98.9 | 0.2 |
| | 242.5 | 239.1 | 98.6 | | |
| | 241.5 | 239.1 | 99.0 | | |
| | 242.1 | 240.0 | 99.2 | | |
| | 244.3 | 241.8 | 99.0 | | |
| | 244.7 | 242.7 | 99.2 | | |
| 120 | 288.8 | 285.6 | 98.9 | 98.9 | 0.2 |
| | 292.5 | 289.3 | 98.9 | | |
| | 290.3 | 287.5 | 99.0 | | |
| | 289.9 | 287.5 | 99.2 | | |
| | 291.7 | 288.4 | 98.9 | | |
| | 291.9 | 287.5 | 98.5 | | |
| Mean (n=18) | | | | 99.0 | 0.3 |

203
204
205
206
207
208
209
210
211
212
213
214




Figure 3. Accuracy

4. CONCLUSIONS

The volumetric method proposed is easy, sensitive and inexpensive and can consequently be applied to the determination of rosuvastatin calcium in raw material. Method validation including linearity, precision and accuracy generated acceptable results.

ACKNOWLEDGMENTS

This work was supported by grant 20020130100342BA to A. I. Segall from UBA and PIP N°: 11420110100380 from CONICET.

REFERENCES

- [1] Beludari MI, Prakash KV, Mohan GK, RP-HPLC method for simultaneous estimation of Rosuvastatin and Ezetimibe from their combination tablet dosage form. *Int J Chem Anal. Sci.* 2013;4:205-209
- [2] Uyar B, Celebier M, Altinoz S, Spectrophotometric determination of rosuvastatin calcium in tablets. *Pharmazie* 2007;62(6):411-413
- [3] Dannana GS, Marothu VK, Extractive Spectrophotometric methods for the determination of Rosuvastatin calcium in pure form and in pharmaceutical formulations by using safranin O and methylene blue. *E. J. Chem.* 2007;4:46-49.

- [4] Gupta A, Mishra P, Shah K, Simple UV Spectrophotometric Determination of Rosuvastatin Calcium in Pure Form and in Pharmaceutical Formulations. *E. J. Chem.* 2009;6(1):89-92.
- [5] Astarita A, DellaGreca M, Iesce MR, Montanaro S, Previtera L, Temussi F, Polycyclic compounds by sunlight exposure of the drug rosuvastatin in water. *J. Photoch. Photobio A.* 2007;187: 263-268.
- [6] Fabris J, Makuc D, Casar Z, Plavec J, Conformational analysis of E/Z-isomeric pairs of rosuvastatin and its lactonized analogues. *Tetrahedron* 2013;69: 6262-6268.
- [7] Litvic M, Smic K, Vinkovic V, Filipan-Litvic M, A study of photodegradation of drug rosuvastatin calcium in solid state and solution under UV and visible light irradiation: The influence of certain dyes as efficient stabilizers. *J. Photoch. Photobio A.* 2013;252:84-92.
- [8] Trivedi HK, Patel MC, Development and Validation of a Stability-Indicating RP-UPLC Method for Determination of Rosuvastatin and Related Substances in Pharmaceutical Dosage Form. *Sci. Pharm.* 2012; 80:393-406.
- [9] Bergman E, Forsell P, Tevell A, Persson AM, Hedeland M, Bondesson U, Knutson L, Lennernäs H, Biliary secretion of rosuvastatin and bile acids in humans during the absorption phase. *Eur. J. Pharm. Sci.* 2006;9:205-214.
- [10] Lan K, Jiang X, Li Y, Wang L, Zhou J, Jiang Q, Ye L, Quantitative determination of rosuvastatin in human plasma by ion pair liquid–liquid extraction using liquid chromatography with electrospray ionization tandem mass spectrometry. *J Pharm Biomed. Anal.* 2007;44:540-546.
- [11] Zhang R, Li Y, Jiang X, Wang L, Pharmacokinetics and Tolerability of Multiple-Dose Rosuvastatin: An Open-Label, Randomized-Sequence, Three-Way Crossover Trial in Healthy Chinese Volunteers. *Curr. Ther. Res.* 2009;70(5):392-404.
- [12] Hauptstein S, Müller C, Dünnhaupt S, Laffleur F, Bernkop-Schnürch A, Preactivated thiomers: Evaluation of gastroretentive minitablets. *Int. J. Pharm.* 2013;456:473-479.
- [13] Balakumar K, Raghavan CV, Selvan NT, Prasad RH, Abdu S, Self-nanoemulsifying drug delivery system (SNEDDS) of Rosuvastatin calcium: Design, formulation, bioavailability and pharmacokinetic evaluation. *Colloid Surface B* 2013;112:337-343.
- [14] Trivedi RK, Kallem RR, Mullangi R, Srinivas NR, Simultaneous determination of rosuvastatin and fenofibric acid in human plasma by LC–MS/MS with electrospray ionization: Assay development, validation and application to a clinical study. *J Pharm Biomed. Anal.* 2005;39:661-669.
- [15] Lee HB, Peart TE, Svoboda ML, Backus S, Occurrence and fate of rosuvastatin, rosuvastatin lactone, and atorvastatin in Canadian sewage and surface water samples. *Chemosphere* 2009; 77:1285-1291.
- [16] Shah Y, Iqbal Z, Ahmad L, Khan A, Khan MI, Nazir S, Nasir F, Simultaneous determination of rosuvastatin and atorvastatin in human serum using RP-HPLC/UV detection: Method development, validation and optimization of various experimental parameters. *J. Chromatogr. B* 2011;879:557-563.
- [17] Nasir F, Iqbal Z, Khan A, Ahmad L, Shah Y, Khan AZ, Khan JA, Khan S, Simultaneous determination of timolol maleate, rosuvastatin calcium and diclofenac sodium in pharmaceuticals and physiological fluids using HPLC-UV. *J. Chromatogr. B*, 2011;879:3434-3443.
- [18] Arayne MS, Sultana N, Tabassum A, RP-LC simultaneous quantitation of co-administered drugs for (non-insulin dependent) diabetic mellitus induced dyslipidemia in active pharmaceutical ingredient, pharmaceutical formulations and human serum with UV-detector *Clin. Chim. Acta* 2013;425:54-61.
- [19] Lee D., Roh H., Son H., Jang S.B., Lee S., Nam S.Y. and Park K., Pharmacokinetic Interaction Between Rosuvastatin and Metformin in Healthy Korean Male Volunteers: A Randomized, Open-label, 3-period, Crossover, Multiple-dose Study. *Clin. Ther.* 2014;36(8):1171-1181
- [20] Roh H, Son H, Lee D, Chan HC, Yu C, Park K, Pharmacokinetic Interaction Between Rosuvastatin and Olmesartan: A Randomized, Open-label, 3-period, Multiple- dose Crossover Study in Healthy Korean Male Subjects. *Clin. Ther.* 2014; 36(8):1159-1170.
- [21] Assassi AL, Roy CE, Perovitch P, Auzeir J, Hamon T, Gaudin K, Green analytical method development for statin analysis *J. Chromatogr. A* 2015; 1380:104-111.
- [22] Narapusetti A, Bethanabhatla SS, Sockalingam A, Repaka N, Saritha V, Simultaneous determination of rosuvastatin and amlodipine in human plasma using tandem mass spectrometry: Application to disposition kinetics. *J. Adv. Res.* 2014, Available from: <http://dx.doi.org/10.1016/j.jare.2014.08.010>
- [23] Skoog DA, West DM, Holler FJ, Química Analítica 6st Ed., Ed. Mc GrawHill, México, 1995
- [24] International Conference on Harmonization. ICH Q2(R1) Guideline on Validation of Analytical Procedures: Text and Methodology (2005).

291 [25] International Conference on Harmonization. ICH Q7 Guideline on Good Manufacturing Practice Guide
292 for Active Pharmaceutical Ingredients (2010).
293
294