



SDI Review Form 1.6

Journal Name:	British Journal of Applied Science & Technology
Manuscript Number:	2014_BJAST_15679
Title of the Manuscript:	RESEARCH OF IMPLANTS COATING STRUCTURE AND DURABILITY
Type of the Article	

General guideline for Peer Review process:

This journal's peer review policy states that **NO** manuscript should be rejected only on the basis of '**lack of Novelty**', provided the manuscript is scientifically robust and technically sound.

To know the complete guideline for Peer Review process, reviewers are requested to visit this link:

(<http://www.sciencedomain.org/page.php?id=sdi-general-editorial-policy#Peer-Review-Guideline>)



SDI Review Form 1.6

PART 1: Review Comments

	Reviewer's comment	Author's comment (if agreed with reviewer, correct the manuscript and highlight that part in the manuscript. It is mandatory that authors should write his/her feedback here)
<u>Compulsory</u> REVISION comments	-	
<u>Minor</u> REVISION comments	Numerical obtained values of sizes of granules are given without incertitude in Table 1, e.g: 50.0 ± 4.2	
<u>Optional/General</u> comments	The authors present an investigation of microstructures even nanostructures: Co+TiO ₂ , Co-Cr-Mo, Zr and ZrO ₂ , used as implant coatings in improved language. The results are obtained, with scanning probe microscopy (SPM) for morphological characterization, before and after the friction test. These results are well analyzed showing that Co+TiO ₂ is the best sample but have to be nevertheless presented according to <i>the rule of thumb</i> .	

Reviewer Details:

Name:	Roland Eba Medjo
Department, University & Country	Physics' Department, Faculty of Science, University of Douala, Cameroon