

Microbiological quality of laboratories works stations: impact of a system of saturated dry spray steam

Running title: Laboratory safety and system of sanitation

Type of article: Original research papers

ABSTRACT

Hygiene and sanitation in laboratories are some important focus for the well-being of scientists and workers. Due to the ~~observed~~ lack of these notions, a new sanitation system named -Polti Sani System has been tested to overcome the limitations of traditional methods ~~about risks of biological contamination. It was the Polti Sani System.~~

For this work, some American Type Culture Collection and Institute Pasteur of Strasbourg reference strains have been used. ~~The biocides activities of dry-saturated steam Nebulizer System which is the Sani System were~~ ~~assessed followed by the count of the total aerobic mesophilic bacteria in the laboratory working environment before and after~~ ~~disinfection.~~ ~~The Adenosine Triphosphate from the different~~ ~~different surface types before and after disinfection was also~~ ~~quantified as well.~~ ~~Based on the results and the experimental conditions adopted, After a time contact of 30 seconds,~~ Sani System reduced more than five logarithmic ~~orders~~ levels, the bacterial rate tested ~~after a time of contact of 30 seconds~~. The log reduction achieved ~~over with~~ the fungi ~~in the same conditions were with the same contact time is greater than~~ greater than four logarithmic ~~orders~~ levels. The antimicrobial activity was observed both in the environment and on inert supports made of glasses, Plexiglas or tiles with an average reduction rate of 99.13%. This study showed that the Sani System associated with ATP-metry can be successfully used to quickly

26 check the hygiene standards on surfaces or lab environment. It is a real challenge in
27 terms of quality, efficiency and safety for the laboratories.

28 **Keywords:** Laboratory safety, sanitation system, saturated steam.

29 1. INTRODUCTION

30 Environment as air, surfaces and water faced a permanent but variable
31 microbiological contamination variable in time and space. Microorganisms belong to
32 the environmental saprophytic floras. It also comes from commensal or pathogenic
33 floras of people [1]. As vectors of contamination, these environmental
34 compartments disseminate microorganisms at some distance and insidiously
35 contribute to progressive contamination of various inert supports.

36 There are many examples of bioburden that have hit the headlines, whether in the
37 hospital, pharmaceutical or food area. Epidemiological data confirm the reality of
38 these risks. Described infections are almost due to bacterial causes and in 13% of
39 cases, it is associated to with exposure to aerosols [2].

40 Causing human and material damage, aerobiocontamination is a fundamental
41 problem which affects many sectors. Biocontamination is a biological contamination
42 which may have an adverse effect on the product (false positives), the staff or the
43 patient with regard to health facilities [2].

44 Regarding Due to this ubiquitous contamination, personnel safety and protection of
45 handled products can only be achieved in a controlled microbial environment [3].

46 Therefore, it is important to assess the risk of contamination and to manage them it
47 appropriately and consistently. The control of bioburden in areas at risk relates to

48 compliance with preventive measures but also on physical, chemical and
49 microbiological tests and finally on correctives measures.

50 The traditional sanitation methods relate to the use of disinfectant solutions,
51 sometimes with high chemical content, which come in contact with the support
52 surfaces or heat or radiation [4].

53 The limitations of these methods are generally related to the difficulty in reaching the
54 interstices or penetrate the rough surfaces and uneven, it does not provide a total
55 cleaning of the contaminated area. In addition, these disinfectants have a risk of
56 irritation or hypersensitivity due to their chemical composition. Moreover, traditional
57 chemical disinfectants are not suitable for all types of surface and the phenomena of
58 natural and acquired resistance to disinfectants were observed in bacteria [5]. In
59 addition, it requires a manual contact with the surfaces to sanitize. Technological
60 advances in microbiology and sanitation, combined with the new requirements of
61 work safety, product safety and environmental protection led to a complete revision
62 of disinfection procedures. Thus, some new products such as aerosols and complex
63 synthetic compounds appeared.

64 Among these procedures, the steam was usually used within controlled enclosures
65 which are capable of withstanding the pressures required for sterilization but its
66 direct use-spray on surfaces is not yet a reality in the world [6]. The principle of such
67 method is particularly interesting because it uses a simple steam and leads to a
68 destruction of microorganisms by the way of generated heat without pollution.

69 Within the overall framework of risk prevention, specifically the reduction of
70 avoidable proportion of airborne infections and also to overcome the resistance
71 problems due to the use of traditional disinfectants, **this study aimed to assess** the

microbiological quality of the lab working environment by using a nebulizer system of dry saturated steam.

2. MATERIALS AND METHODS

2.1. Materials

Materials used were composed of microbial reference strains, experimental equipment disinfection system (nebulization of saturated steam Sani System) luminometer, culture media and reagents, glasswares and consumables.

2.1.1. Microbial strains reference

The antibacterial and antifungal activities of the disinfectant ~~was~~ were *in vitro* evaluated on ATCC (American Type Culture Collection) and IP (Institute Pasteur of Strasbourg) strains ~~namely: of~~ *Enterococcus hirae* ATCC 10541, *Staphylococcus aureus* ATCC 29213, *Escherichia coli* ATCC 25922, *Pseudomonas aeruginosa* ATCC 27853, *Candida albicans* IP 4872 and *Aspergillus Niger* ATCC 16404.

2.1.2. Experimental disinfection equipment

Disinfection equipment distributed by the company MIVA was a fogging system saturated vapour and a sanitizer. Sani System is an electro-medical device which delivers saturated steam at high temperature (180°C) associated with the sanitizing HPMed. Its uniqueness lies in the fact that the steam is brought to a high temperature in an expansion chamber (Figure 1).



94
95
96
97
98
99

Figure 1. A nebulizer with dry saturated steam: Sani System Polti

100 HPMed is a co-adjuvant in sanitation vapour of Sani System. It consists of an
101 alcohol-based solution of sodium metasilicate containing a nonionic surfactant, which
102 helps the sanitizing activity of Sani System. The maximum flow rate of the saturated
103 steam is 100 g/ min HPMed represents an aid to sanitation carried out by
104 superheated dry saturated steam delivered by the Polti Sani System. The
105 guaranteed minimum consumption HPMed is 0.4 ml / min Table I shows the
106 chemical composition of the sanitizer HPMed.

107 **Table 1:** Composition of disinfectant HPMed

Ingredients	Metasilicate sodium	Carbonate sodium	Ethoxylate alcohol	Ethyl alcohol
Percentage	0.1-1%	0.1-1%	0.1-1%	5-10%

108 NB: Composition: Nonionic surfactant <5%

109 A 3 M luminometer clean-trace NG was used with 3 M kits reagents for measuring
110 contamination levels in samples of surface and can effectively monitor surface
111 hygiene. The reagent kits were some dry swabs and Lucifer in-luciferase complex
112 [20].

113 **2.2. Methods**

The methodology used was about the evaluation of the biocidal activities of the nebulization Sani System. Then, the numeration of total aerobic mesophilic germs in a working environment (air and hard surfaces) before and after disinfection was done. Finally, the Adenosine Triphosphate from the different types of surface was quantified before and after disinfection.

2.2.1. Evaluation of biocides activities of Sani System Polti

The preparation of the microbial suspension was made according to AFNOR NF T 72-281 [7] where the dilution which gives a concentration of 10^6 germs per ml was selected as microbial suspension. Biocidal activities of Sani System Polti have been assessed by the test support [8, 9]. After contamination and drying, some slides were treated by spraying of dry saturated steam in combination with sanitizer HPMed for 30 seconds. Along with these tests, another slides contaminated by each microorganism and untreated were maintained at room temperature (16°C) throughout the tests for determining the initial contamination level brackets and validation of experimental conditions.

After treatment by the Sani System, the slides (N' for the test and N for the untreated control) were transferred into 10 mL of neutralizer (polysorbate 80: 50 mL, sodium thiosulfate: 5 g, trypticase soy broth: qsp 500 mL). The erlenmeyers containing the test slides were then subjected to stirring for 30 seconds and the number of viable microorganisms per mL of reaction medium was determined by successive decimal dilutions following by plating of each dilution on Trypticase Soy Agar (TSA). Each microorganism was tested three times. Seeded Petri dishes and controls were

incubated under appropriate conditions for each microorganism. Tryptic Soy Agar was incubated at 35 ± 2 °C for 48 hrs in bacteriological incubator Memmert IBN400. Sabouraud agar with chloramphenicol was incubated at 25 ± 2 °C for 96 hrs in incubator Memmert IPP 400. After incubation, the colonies formed were counted using colony counter Stuart Scientific and the results were expressed in CFU / cm².

For each strain, the log decimal reduction between the number of microorganisms present on the slide before treatment (N) and after treatment (N') was calculated using the formula:

$$R = N/N'; \text{ Where } \log_{10} R = \log_{10} N - \log_{10} N'$$

2.2.2. Numeration of mesophilic aerobic microorganisms suspended in the air

Mesophilic aerobic bacteria were collected and counted in the environment of three rooms before and after disinfection for respectively 360 seconds, 600 seconds and 3600 seconds. The method used was the sedimentation one on Petri dish [10]. Three samples were taken in each room and each sample was assayed five times. The time between two (02) samples was seven days. After cultivation staggered between 30-35 °C for 48 hrs and 20-25 °C for 96 hrs, the result of numeration R was expressed as follows:

$$R = \frac{(N_1+N_2+N_3+N_4+N_5)}{5} \times 2h$$

R is expressed in CFU/4 hrs.

N₁, N₂, N₃, N₄, N₅ are the number of colonies counted for each of the five (5) plates seeded.

2.2.3. Enumeration of total aerobic mesophilic floras from lab surfaces

The samples were collected on three types of surfaces: glass, plexiglas and tile (before and after 30 seconds of disinfection) using the technique of wet swab on an area of 10 cm². Ten samples were taken from the surface and 0.1 mL of each dilution sample was plated in duplicate using Tryptic Soy Agar [11]. After incubation at 30-35 °C for 72 hrs, the results expressed as CFU / ml of the initial product, were converted to counts per cm² using the following formula [12]:

$$N' = N \times 4 \text{ where:}$$

N': Number of CFU per cm².

N: Number of CFU per ml of the original product.

2.2.4. ATP-metry: measurement of residual ATP before and after disinfection

Surface samples were collected (before and after 30 sec of disinfection) with dried swabs (ATP test surface). Light emitted by the ATP test surface was measured and the result was displayed on the screen of the luminometer.

2.2.5. Statistical analysis

The results were analysed regarding the disinfection time and the type of surface. To evaluate the influence of the duration of disinfection and the type of surface on the rate of reduction of microorganisms, the test of analysis of variance was used when the conditions of normality and equal variances were met. Otherwise, the Kruskal Wallis was preferred. Student's t test or Wilcoxon test was used to compare the rates of reduction of microorganisms obtained for each factor. The significance level was set at 5%.

3. RESULTS AND DISCUSSION

3.1. Effectiveness of Sani System Polti

The table 2 presents the results of the effectiveness of Sani System Polti on microorganisms. It shows that the supports contaminated with different bacterial strains and treated with Sani System have no detectable microbial growth. The log reductions highlighted after 30 sec disinfection with dry saturated steam were all higher than those required in the NF EN 14561: 2007, EN 14562: 2006 standards [13, 14].

Table 2: Biocidal activities of HPMed sanitizer

	<i>E. hirae</i>	<i>S. aureus</i>	<i>E. coli</i>	<i>P. aeruginosa</i>	<i>C. albicans</i>	<i>A. niger</i>
Untreated Supports (N)	1.2 x10 ⁶	1.5 x10 ⁶	1.4x 10 ⁶	1.2 x10 ⁶	1.0 x10 ⁶	1.1 x10 ⁶
Treated Supports (N')	<1	<1	<1	<1	20	30
log reduction (R)	> 6.07	> 6.17	> 6.14	> 6.07	4.69	4.56

These results confirm those of [6] who have also obtained more than 5 log units reductions for bacteria (*Pseudomonas aeruginosa*, *Staphylococcus aureus*, *Enterococcus hirae*) and reductions greater than 4 log units for *Aspergillus niger* and *Candida albicans*. Meanwhile, the log reductions obtained in this study for fungi differ from those reported by [15] who highlighted, for a similar system, some reductions discounts greater than 5 log units for *Candida albicans*.

3.2. Numeration of mesophilic aerobic microorganisms in the air

200 The floras for the three rooms (S_1 , S_2 and S_3) before and after disinfection are
 201 mentioned in the **table 3**. More disinfection time increases, higher is the reduction of
 202 the number of microorganisms.

203

204

205

206 **Table 3:** Presence of mesophilic aerobic microorganisms in the three rooms

		R_1 (CFU/4h)			R_2 (CFU/4h)			R_3 (CFU/4h)		
		J1	J8	J15	J1	J8	J15	J1	J8	J15
After 360 sec	Not disinfected rooms	28	24	34	36	42	48	44	52	58
	Disinfected rooms	26	24	32	32	40	46	42	52	56
	% reduction	7.14	0	5.88	11.11	4.76	4.16	4.54	0	3.44
After 600 sec	Not disinfected rooms	36	32	44	66	58	64	80	86	84
	Disinfected rooms	18	18	24	34	28	34	58	60	60
	% reduction	50	43.75	45.45	48.48	51.72	46.87	27.5	30.23	28.57
After 3600 sec	Not disinfected rooms	34	38	36	74	70	68	88	92	90
	Disinfected rooms	2	2	2	4	4	4	6	6	4
	% reduction	94.11	94.73	94.44	94.59	94.28	94.11	93.18	93.47	95.55

207

208 Microbial quantum in the three rooms (R_1 , R_2 , and R_3) after disinfection during 3600
 209 seconds is respectively 02, 04 and 05 CFU/4 **hrs**. All these rooms belong to Class B
 210 as recommended by limits about microbiological **contamination [16]**. The average

reduction of airborne microorganisms in three rooms after 360 sec, 600 sec and 3600 sec nebulization is shown in Figure 2. It shows that the rate of reduction of airborne germs in three rooms increases with duration of disinfection. Thus, the average reduction rate was 4.57% after 360 sec disinfection. This rate raised to 41.39% after treatment of 600 sec and reached 94.27% after 3600 sec of disinfection.




Figure 2: Evolution of the average rate of reduction of microorganisms by nebulization time

The results of the analysis of variance of two factors are presented in the Table 4_[WU1]. The duration of disinfection has a significant effect ($p\text{-value} < 0.05$) on the rate of reduction of microorganisms. This time would probably depend on the surface to be disinfected. The analysis of variance of two factors also indicates that the disinfecting time duration has a highly significant effect on the rate of reduction of airborne germs. Although, disinfection is a one-time treatment effect and does not persist over time ($p\text{-value} = 0.79$).

236

Table 4: Analysis of variance of two factors

	Degree of Liberty	Sums of squares	Mean squares	F statistic	p-value
Nebulization Time	2	36594	18297	476.5011	<0.001
time	2	18	9.1	0.2367	0.79

237

238

239 3.3. Numeration of mesophilic aerobic microorganisms on surfaces

240 The boxplot in Figure 3 shows the distribution of the average rate of reduction of
 241 microorganisms by type of surfaces. The average rate of reduction of
 242 microorganisms on glass and plexiglas has a wide variation different from that of the
 243 tiles. The reduction rate corresponding to the average numbers were respectively
 244 98.92 %, 98.51 % and 99.98 %. The system with saturated steam is effective on
 245 smooth surfaces. These results are similar to those of [16] who obtained a rate
 246 reduction between 98-100 % for smooth surfaces. Moreover, none of the treated
 247 surfaces have been damaged. No fundamental change in colour, shape and general
 248 appearance has been noted.

249

250

251

252

253

254

255

256

257




Figure 3: Distribution of the average rate of reduction of microorganisms by type of surface

Kruskal Wallis test shows that the nature of the surface significantly influences the rate of reduction of microorganisms after disinfection. Indeed, the rate of reduction of microorganisms after nebulization of ground (tiles) is significantly higher ($p < 0.05$) than the rate of reduction of the smooth surfaces.

This result is contrary to the theory that the use of traditional methods of disinfection (towel) on surfaces of low roughness (shiny and well polished) provides a high rate of reduction [18]. The difference should be partly due to the nature of the microorganisms which are very sensitive to heat and sanitizer on the tiles. On the other hand, the fact that the vapours, unlike napkins, penetrate anywhere, even areas closer could explain such situation. By cons, there is no statistical difference between the rate of reduction of a glass surface and a Plexiglas surface ($p = 0.71$).

3.4. ATP-metry

The average of different values taken by the luminometer before and after disinfection of different types of surfaces (tiles, glass and Plexiglas) and the average reduction are shown in Table 5. The average rate of reduction of the relative amount of light into surfaces of tiles is 94.23% against 93.85% for the glass surfaces and 90.56% for plexiglas. The average of RLUs for glass surfaces, plexiglas and tiles were respectively 37.7 RLU/cm², 68.4 RLU/cm² and 171.5 RLU/cm². These values are well below the threshold of acceptability in terms of safety (0-250 RLU/cm²) [19].

The rate of disinfection on glass and tile is significantly higher than the rate of disinfection in the plexiglas ($p < 0.05$). However, there is no statistically significant difference between the rate of reduction of microorganisms on the tile and glass ($p = 0.73$). Although, whatever the disinfected surface, the reduction rate estimated by ATP-metry method was significant ($p < 0.05$).

Table 5: Average rate reduction RLUs by area

Surfaces	before disinfection RLU	after disinfection RLU	Average reduction RLU _s % (sd*)	p-value
Tiles	2437.5	171.5	94.23 (3.17)	<0.05
Glass	603.8	37.7	93.85 (1.40)	<0.05
Plexiglass	719.9	68.4	90.56 (1.13)	<0.05

* : sd = standard deviation

It is often wrongly assumed that the results obtained by Relative Light Unit (RLU) should demonstrate a direct correlation with the numeration on petri dish for the same sample. The amount of ATP present in microbial cells can vary greatly depending on the strain and one CFU may correspond to one or more microorganisms. In addition, the bacteria need to grow with food waste so even if in theory, it is possible to make a correlation between RLU and CFU, in practice, however it is impossible to determine the exact origin of ATP (bacteria, yeast, mold, food residue). The most frequent case is a combination of multiple factors. Thus the ATP-metry is not a direct indicator of the presence of bacteria. Nevertheless, it is the

fastest and easiest technology to determine the potential for bacterial growth. Therefore, it is an indicator of the actual surface level of hygiene. There is no easier or faster else way to determine a problem. Thus, with the ATP-metry, it is possible to implement corrective actions upstream to avoid any problem of non-compliance.

CONCLUSION

Microbial contamination is a serious concern for many industries including food factories, laboratories as well as hospital centres whereby more research are needed on the aspect of disinfection for safety purposes. Methods that can reduce or eliminate microbial growth with less or no injury for the operator and the environment are preferred. This is the case of disinfection using dry saturated steam Sani System which presents many advantages for sanitation despite the cost. In view of these advantages, it is quite possible to consider other applications in the food and medical areas where food and nosocomial infections are not negligible. Associated with ATP-metry, Sani System is then a real challenge in terms of quality, efficiency and safety.

CONFLICT OF INTERESTS

The author(s) declare(s) that there is no conflict of interest regarding the publication of this paper.

REFERENCES

1. Squinazi F. Mastering the microbial environment in the hospital about the aerobio-contamination - Heating, ventilation, air conditioning. Hospital technology 1995/03; (595): 51-56. French.

- 324 **2.** Pike RM. Laboratory associated infection: Incidence, fatalities, cases and
325 prevention. Annual Review of Microbiology, Ann. Rev. Microbiol. 1979;
326 (33): 41-66.
- 327 **3.** Good Hospital Pharmacy Practices. First edition. 2001; 72p. French.
- 328 **4.** Fleming DO, Hunt DL. Biological Safety: principles and practices 3rd edition.
329 Washington DC: ASM press. 2000; 784p.
- 330 **5.** Chopra I. Bacterial resistance to disinfectants, antiseptics, and toxic metal
331 ions. In Mechanisms of action of chemical biocides, their study and
332 exploitation (S. Denyer & W.B. Hugo, edit.). Society for Applied Bacteriology,
333 Technical Series N°. 27. Blackwell Scientific Publications, Oxford. 1990; 45-
334 64.
- 335 **6.** Pineau L, Desbuquois C. Disinfection by steam: microbiological efficacy.
336 Hygiene. 2007; 15 (4): 7p. French.
- 337 **7.** French Association for Standardization. AASD processes, determining the
338 bactericidal, fungicidal, and sporicidal yeasticidal: French NF T. 2009; 72-281.
- 339 **8.** Cremieus A, Fleurette J. Methods of testing disinfectants. In disinfection,
340 sterilization, and preservation, 4th ed. SS.block, London. 1991;
341 1009-1021.
- 342 **9.** Lambert RJW. Evaluation of antimicrobial efficacy. In: Principles and practice
343 of disinfection, preservation and sterilization (fourth edition). AP. Fraise,
344 Lambert, PA., Maillard, JY. (Eds). Etats Unis. 2004; 345-360.

10. Morey P, Otten J, Burge H, Chatigny M, Feeley J, LaForce FM, Peterson K. Airborne viable microorganisms in office environments: sampling protocol and analytical procedures. *Applied Industrial Hygiene*. 1986; **1**:19-23.
11. Guyader P. Amgar A, M. Coignard Disinfection: verification and validation of the effectiveness of the disinfection operations. In: Bourgeois, CM, Mescle, JF, Zucca, J Coord. Tome 1 Food Microbiology: Microbiological aspect of safety and quality of food. Paris: Lavoisier Tec & Doc. 1996; pp. 451-455. (Technical and food sciences).
12. Bendeddouche B. and Bensid A.: Monitoring the effectiveness of cleaning and disinfection of equipment in a poultry slaughterhouse in Algeria. *European Journal of Scientific Research*. 2009; 27(2): 181-187. French.
13. NF EN 14561. Chemical disinfectants and antiseptics - Quantitative carrier test for the evaluation of bactericidal activity for instruments in medicine - Test method and requirements (phase 2, step 2). 2007.
14. NF EN 14562. Chemical disinfectants - Quantitative carrier test for the evaluation of fungicidal or yeasticidal activity for instruments used in human medicine - Test method and requirements (phase 2, step 2). 2006.
15. Haas A, Platz S, Eichhorn W. Effect of steam application based on microbiological and parasitologic test procedures. *Zentralbl Hyg Umweltmed*. 1998; 201(4-5): 337-347.
16. European Union. EU Good Manufacturing Practice. Medicinal Products for Human and Veterinary Use. Annex 1 Manufacture of Sterile Medicinal

367 Products (corrected version). 91/356/EEC, EudraLex - Good manufacturing
368 practice (GMP). 2008; (4).

369 **17. Carcano E.** Evaluation of the kinetics of the bactericidal effectiveness of Polti
370 Sani System: quantitative microbiological testing by means of total aerobic
371 bacteria count Italie. 2008; 12p.

372 **18. Bourrion F.** Limitations of cleaning and disinfection operations: biofilms. In:
373 Albert A. Coord. Cleaning and disinfection in food companies. Laval: ASEPT.
374 1998; pp. 205-211.

375 **19. Champiat D.** Applications of bioluminescence to HACCP.
376 2001;16(2):193–198.

377 **20. Champiat D., & JP Larpent JP.** Bioluminescence: principles and
378 applications. Masson biotechnology; 1993- 531p. French

379

380

381

382

383

384

385

386

387

388