

Editorial comments:

One of the reviewers (MS) pointed out that the refractometer may give false information about the NaCl concentration, particularly if there are other ions in the water sample. It seems to me that the refractometer only reflects salinity of the water. The authors should consider changing the NaCl concentration to salinity throughout the manuscript.

Author's feedback:

Thank you for comment. Ms corrected. In attached we have not "NaCl concentration".